

BioSim™ anti-Daratumumab (Human) ELISA Kit

(Catalog # E4924-100, 96 assays, Store at 4°C)

I. Introduction:

Daratumumab is a human IgG1 monoclonal antibody that binds with a high affinity to a unique epitope of CD38. CD38 is a type II transmembrane glycoprotein expressed on lymphoid, non-hematopoietic tissues, and myeloid cells. CD38 possesses ectoenzyme activity and plays a significant role in receptor-mediated regulation of cell adhesion and signal transduction. It is expressed at high levels on multiple myeloma cells, thus making it a unique target for therapeutic antibodies targeting multiple myeloma cell surfaces. The monoclonal antibody binds to CD38 on the surface of multiple myeloma cells and induces cell death via multiple mechanisms: complement-dependent cytotoxicity (CDC), antibody-dependent cell-mediated cytotoxicity (ADCC), antibody-dependent cellular phagocytosis (ADCP), and apoptosis. The original antibody is approved to treat multiple myeloma in patients who have received at least 3 previous therapies. However, upon administration of the antibody, patients tend to develop an immune response that results in the formation of anti-Daratumumab antibodies, thus adversely affecting the therapeutic response to treatment. BioSim™ anti-Daratumumab ELISA kit qualitatively measures antibodies to Daratumumab in human serum or plasma samples. The kit is based on the Sandwich ELISA principle. Standards and samples (serum or plasma) are added to the microtiter plate coated with Daratumumab. After incubation, the wells are washed. The HRP conjugated probe is added and binds to the antibodies to Daratumumab that is captured by Daratumumab coated on the wells. Following incubation, wells are washed and the enzymatic activity is detected by the addition of TMB chromogen substrate. Finally, the reaction is terminated with an acidic stop solution. The color developed is proportional to the amount of antibodies to Daratumumab in the sample or standard. The qualitative test results can be evaluated using cut-off values.

II. Features and Benefits:

- For *in vitro* Qualitative determination of antibodies to Daratumumab in human serum and plasma samples
- Assay Precision: Intra-Assay and Inter-Assay CV < 30%
- Cross Reactivity: Except for Daratumumab, there is no cross reaction with other therapeutic antibodies and native serum immunoglobins

III. Sample Type:

Human Plasma and Serum

IV. Kit Contents:

Components	E4924-100	Part Number
Microtiter Plate	8 X 12 Strips	E4924-100-1
Positive Control	0.3 ml	E4924-100-2
Negative Control	1 ml	E4924-100-3
Assay Buffer	12 ml	E4924-100-4
HRP-conjugate Probe	12 ml	E4924-100-5
TMB substrate (Avoid light)	12 ml	E4924-100-6
Stop Solution	12 ml	E4924-100-7
Wash buffer (20X)	50 ml	E4924-100-8
Plate sealers	2	E4924-100-9

V. User Supplied Reagents and Equipment:

- Microplate reader capable of measuring absorbance at 450 nm
- Precision pipettes with disposable tips
- Clean eppendorf tubes for preparing standards and sample dilutions
- Absorbent paper

VI. Storage and Handling:

The entire kit may be stored at 4°C for up to 12 months from the date of shipment

VII. Reagent and Sample Preparation:

Note: Before using the kit, spin the tubes and bring down all the components to the bottom of the tubes

1. **Wash Buffer:** Dilute the 20X Wash Buffer to 1X solution using ddH₂O (10 ml of 20X Wash Buffer + 190 ml of ddH₂O). Mix the 1X solution thoroughly by manual vortexing. The working stock is stable for 2 weeks after preparation at 4°C. If there are crystals in 20X Wash Buffer, warm the Buffer at 37°C prior to dilution.
2. **Sample preparation:** The usual precautions should be observed for venipuncture. Samples that are hemolytic, icteric or lipemic should be avoided. If the sample is turbid, then it must be centrifuged to separate particulates from solution. Freeze/thawing of serum/plasma samples should be avoided. Drug infusions may interfere with the detection of antibodies to drugs in serum/plasma samples. Hence, it is advisable to take blood samples prior to the scheduled dose. Collected samples are stable for 2 days at 4°C or for 6 months at -20°C.

VIII. Assay Protocol:

Note: Bring all the reagents, samples and microtiter plate to room temperature 15 minutes prior to the assay.

It is recommended that all samples be run at least in duplicates.

FOR RESEARCH USE ONLY!

1. Prepare all reagents and samples as instructed in section VII.
2. Pipette 100 μ l of **Assay Buffer** into each of the wells to be used.
3. Add 10 μ l of **Negative control** (2 wells), **Positive control** (1 well), and **samples** into appropriate wells. Cover the wells with Plate sealer and incubate at room temperature (RT) for 60 minutes.
4. Remove the Plate sealer and discard the incubation solution. Wash plate 3 times with 300 μ l of 1X **Wash Buffer**. Remove excess solution by tapping the inverted plate on an absorbent paper.
5. Add 100 μ l of **HRP-conjugated probe** into each well. Cover the plate with adhesive plate sealer and incubate at RT for 60 minutes.
6. Discard the solution and wash the wells as in step 4.
7. Add 100 μ l of **TMB substrate** solution and incubate the plate in dark at RT for 20 minutes.
8. Add 100 μ l of **Stop solution** to stop the reaction. Color changes from blue to yellow.
9. Read the absorbance in micro plate reader set to 450 nm within 30 minutes after pipetting the **Stop solution**. (Use reference wavelength as 650 nm)

IX. Interpretation of Results:

For the run to be valid, the OD 450/650 nm of the Positive control should be > 1.500 and the OD 450/650 nm of each Negative control should be < 0.150. If the results do not comply with the aforementioned information, then improper technique or reagent deterioration may be suspected and therefore the assay must be repeated.

The results are evaluated by a cut-off value which is estimated by multiplying the mean OD 450/650 nm of the negative controls by 3

- If "Sample OD450/650 / the mean Negative Control OD450/650" is < 3, the sample is NEGATIVE for antibodies to Daratumumab.
- If "Sample OD450/650 / the mean Negative Control OD450/650" is \geq 3, the sample is POSITIVE for antibodies to Daratumumab.

Note: The cut-off information provided with this kit can only be considered as a recommendation. Cut-off values must be calculated/set or verified according to scientific standards by the users.

X. Related Products:

- BioSim™ Rituximab (Human) ELISA Kit (Cat. No. E4371-100)
- BioSim™ Adalimumab (Human) ELISA Kit (Cat. No. E4372-100)
- BioSim™ Bevacizumab (Human) ELISA Kit (Cat. No. E4373-100)
- BioSim™ Etanercept (Human) ELISA Kit (Cat. No. E4374-100)
- BioSim™ Infliximab (Human) ELISA Kit (Cat. No. E4375-100)
- BioSim™ anti-HER2 (Human) ELISA Kit (Cat. No. E4376-100)
- BioSim™ Golimumab (Human) ELISA Kit (Cat. No. E4377-100)
- BioSim™ anti-HER2 mab (Human) ELISA Kit (E4386)
- BioSim™ Cetuximab (Human) ELISA Kit (Cat. No. E4379-100)
- BioSim™ Denosumab (Human) ELISA Kit (Cat. No. E4380-100)
- BioSim™ anti-Denosumab (Human) ELISA Kit (E4394)
- BioSim™ Nivolumab (Human) ELISA Kit (Cat. No. E4382-100)
- BioSim™ Pembrolizumab (Human) ELISA Kit (Cat. No. E4383-100)
- BioSim™ Avelumab (Human) ELISA Kit (Cat. No. E4384-100)
- BioSim™ Durvalumab (Human) ELISA Kit (E4925)
- BioSim™ anti-Durvalumab (Human) ELISA Kit (E4926)
- BioSim™ anti-Pembrolizumab (Human) ELISA Kit (E4397)
- BioSim™ anti-Nivolumab (Human) ELISA Kit (E4396)